



America

CERTIFICATE

No. U8 16 02 12164 011

Holder of Certificate: **Sedlbauer AG**
 Wilhelm-Sedlbauer-Str. 2
 94481 Grafenau
 GERMANY

Production Facility(ies): 12164

Certification Mark:



Product: **Medical power supplies
 Isolations transformer**

Model(s): **MTT 1000**

Parameters:

Rated voltage:	100V / 120V / 230V AC
Rated current:	9,1A / 8.9A / 4.7A
Rated frequency:	50/60 Hz
Protection class:	I

Tested according to: ANSI/AAMIES 60601-1:2005/(R)2012
 CAN/CSA-22.2 No. 60601-1:14

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 713073084_MTT

Date, 2016-02-02

Page 1 of 1

