



America

CERTIFICATE

No. U8 16 07 12164 013

Holder of Certificate: Sedlbauer AG
Wilhelm-Sedlbauer-Str. 2
94481 Grafenau
GERMANY

Production Facility(ies): 91164

Certification Mark:



Product: Medical power supplies
Isolations transformer

Model(s): polyMIT® 350, polyMIT® 350 e
polyMIT® 700, polyMIT® 700 e
polyMIT® 1000, polyMIT® 1000 e

Parameters:

Rated voltage:	115/230 V
Rated frequency:	50/60 Hz
Protection class:	I
Rated power:	350 VA; 700 VA; 1000 VA

Tested according to: ANSI/AAMI ES 60601-1:2005/(R)2012
CAN/CSA-22.2 No. 60601-1:14

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 713085906_polyMIT

Date, 2016-08-05

Page 1 of 1

